



SDI Review Form 1.6

Journal Name:	International Research Journal of Pure and Applied Chemistry
Manuscript Number:	2013_IRJPAC_6440
Title of the Manuscript:	Reverse Flow Injection–Spectrophotometric Determination of Ethylene Glycol via Oxidative Reaction
Type of the Article	

General guideline for Peer Review process:

This journal's peer review policy states that **NO** manuscript should be rejected only on the basis of '**lack of Novelty**', provided the manuscript is scientifically robust and technically sound.

To know the complete guideline for Peer Review process, reviewers are requested to visit this link:

(<http://www.sciencedomain.org/page.php?id=sdi-general-editorial-policy#Peer-Review-Guideline>)



SDI Review Form 1.6

PART 1: Review Comments

	Reviewer's comment	Author's comment (if agreed with reviewer, correct the manuscript and highlight that part in the manuscript. It is mandatory that authors should write his/her feedback here)
<u>Compulsory</u> REVISION comments	Include a comparison table for the reference 9-17. Include a paragraph on FIA in introduction and include the following references. Analytical Applications of Flow Injection Chemiluminescence for the Determination of Pharmaceuticals—A Review. Current Pharmaceutical Analysis. 9(4) (2013), 363-395. Flow Injection methods for the determination of α-tocopherol with spectrophotometric detection. Journal of The Chemical Society of Pakistan, 33(4) (2011) 508–514. Flow injection spectrophotometric determination of vitamin E in pharmaceuticals, milk powder & blood serum using potassium ferricyanide-Fe(III) detection system. Chinese Chemical Letters, 21(6) (2010) 712–715. Why the absorbance is given in mV units in figures 5-12???	A comparison table was added. The paragraph was added and it includes the suggested references. Because the absorbance values measured from recorder with (mV) unit.
<u>Minor</u> REVISION comments		
<u>Optional/General</u> comments		