1 2 Original Research Article 3 Development of Low-cost Viscosity 4 Measuring Device using Vibrational 5 Technique

11 ABSTRACT

The project centers on the development of viscosity measuring device using vibrational techniques. The instrument consists of a variable power supply, two solenoids each with plungers, needle with a pendulum bob attached, switching circuit, current – voltage sensors, microcontroller, analog-to-digital converter and Liquid Crystal Display (LCD) unit. The developed system consists of a variable power supply used to create varying magnetic fields in the solenoids so that the plungers can attract and release the pendulum bob. As the magnetic field strength varies, the pendulum experiences force sufficient to overcome the internal friction within the liquid. The pendulum bob is thus set into periodic to and fro motion. The pendulum bob movement was monitored by a microcontroller through a reflective optical sensor. Viscosity was measured through the calibration constant generated from a liquid of known viscosity. applied voltage, current generated from the solenoids and the dragging time. Gear oil, engine oil, palm oil, olive oil were the fluids investigated with the developed system for performance evaluation. Statistical analysis revealed a percentage error value of 2.87% and accuracy of 97.13% which shows a good agreement between the standard and what is available in literature. The instrument performed well and it is therefore suitable for the measurement of liquids' viscosities.

12 13 14 15

16

Keywords: Magnetic field strength, Calibration constant, Solenoid, Microcontroller

1.0 INTRODUCTION

17 18 The concept of viscosity was postulated by Sir Isaac Newton as the resistance of a fluid (liquid or gas) to a change in 19 shape, or movement of neighboring portions relative to another and denotes opposition to flow. Viscosity is also 20 referred to as lack of slipperiness [1]. Using this concept, all fluids can be categorized by examining the relation 21 between the applied shear stresses and the deformation rate of flow of a fluid [2].

Viscosity is the quantitative property of a fluid, be it liquid or gas, and can also be used in quality control [3]. A change in viscosity can indicate a fundamental change in a material under test [4]. Viscosity measurement and control has great importance in food industry and an accurate knowledge of viscosity is necessary for various industrial processes [5]. Accurate viscosity data and their variations with operating conditions are needed also for various research and engineering applications in any branch of the food industry, such as developing food processes and processing equipment, control of products, quality evaluation and an understanding of the structure of food and raw agricultural materials [6].

- Also, the knowledge of viscosity is of primary importance to the fruit industry. For example, viscosity is an important
- 30 property of juices in the production of heat- and mass-transfer coefficients and in the design of heat-and mass-transfer
- 31 equipment for juice industry. It is also of great importance in automobile industry because the knowledge of viscosity
- 32 in engine oils is used as a method of maintaining the performance of engines.
- 33 So many researches have been conducted on viscometry but none of them has addressed variation of liquid samples.
- The theory of the vibrating wire viscometer was improved by the work of Mostert *et al* in 1989 [7]. Their publication includes the equations necessary to calculate viscosity based on the mechanical motion of wire and the motion of the
- surrounding fluid. In 2005, Kandil made some modifications by varying the material, radius, length of wire used in the
 measurement of gases and liquids viscosities. He achieved a better accuracy with an estimated uncertainty of about
 2% [8].
- A rod of much larger diameter instead of a wire was experimented by Retsina in 1987. His model was tested on water and it produced an accuracy of 0.1% [9]. A vibrating wire viscometer requires a smaller sample fluid volume than the falling body and oscillating viscometers. Moreover, its construction is also very simple relative to the other methods, securing the reproducibility of results.
- Furthermore, the technique of viscosity measurement made by Won-kyu *et al.*, involved a small sample of material levitated and melted in a high vacuum using a high temperature electrostatic levitator [10]. The resonant oscillation of the drop was induced by applying a low level electric field pulse at the drop of resonance frequency; the transient signals which followed the pulses were recorded and the viscosity signal was then extracted. This technique is useful for both molten metallic alloy and semiconductors. The system is mostly designed for high temperature liquid.
- 47 for both molten metallic alloy and semiconductors. The system is mostly designed for high temperature liqu
- This work makes use of a vibrational system produced by varying magnetic field to drag a pendulum bob within the liquid sample under test. It measures the dragging time in relation with the applied voltage across the solenoid; and then compute the viscosity of the liquid sample as decribed below.

52 2.0 MATERIAL AND METHODS

53 54 Figure 1.0 shows the block diagram of the developed instrument for automatic measurement of liquid's viscosity. The 55 instrument consist mainly of a vibrational unit, optical sensor, microcontroller unit, and a display unit. The vibrational unit is made up of two solenoids with plungers and a pendulum bob suspended vertically by a thin rod (needle) placed 56 between the common ends of the two plungers as shown in figure 2. The variable power supply provides alternating 57 58 magnetic fields in the two solenoids which causes the plungers to push and pull the needle which holds the pendulum 59 bob. Hence the to-and fro movement of the pendulum bob was provided as the solenoids get energized and de-60 energized gradually. The optical sensing unit is made up of an optocoupler placed adjacent to the transparent liquid 61 container for sensing the to and fro movement of the bob. The optocoupler is embedded with a Light Emitting Diode 62 (LED) which offers the transmitting function and the phototransistor which serves as the receiver. The optocoupler 63 utilized for this project was PS4009 and it was made to work in a switching mode. The optocoupler detects the to and 64 fro movement of the bob as the bob intercepts the ray path between the sender (LED) and receiver (phototransistor) 65 during oscillation. The period of oscillation is an analogue parameter while data acceptable by microcontroller must be 66 in digital form. This period (time) is an example of physical quantity existing in analogue form. To convert a 67 continuously varying physical guantity into electrical signals, there is need for an Analog-to-Digital Converter (ADC) to 68 translate the analog signals to digital numbers so that the signal can be recognised by microcontroller [11]. 69



51

Figure 1: Block diagram of the developed viscometer

71



85 86

87 88

Figure 2: Image of the developed viscometer

The microcontroller also function as voltmeter and ammeter by measuring the applied voltage and current required to overcome the liquid's internal friction and set the bob into motion. The microcontroller unit consists of a programmable integrated circuit known as atmega 328P. A suitable micro-C code was written and embedded in the microcontroller so as coordinate the activity of the entire system, perform the necessary calculations and send output result to the 83 display unit. The microcontroller was interfaced with a 2 by 16 Hitachi Liquid Crystal Display (LCD) so as to display 84 the measured values of current, voltage, time and viscosity.

2.1 **MODE OF OPERATION**

Figure 3.0 shows operational chart of the developed instrument. This work made use of a linear type solenoid 89 90 because of the need for the plunger to move in and out when variable is power fed into the solenoids from the 91 microcontroller. A 15 V adjustable power supply was used to provide digital voltage ranging from 1.2 V to 15 V to 92 power the MCU, transistors and optocoupler. The alternating output voltage from the MCU was amplified by 93 transistors. The outputs of these transistors were connected to the commonly open and normal terminals of the relay 94 so as to create varying magnetic field in the two solenoids one after the other. When the solenoid produces magnetic 95 force which is just sufficient to overcome the viscosity of the liquid, the pendulum bob begins to move to-and-fro. The 96 reflective optocoupler positioned directly adjacent to the transparent liquid container senses the ON and OFF states of 97 the solenoids as the pendulum oscillates.



Figure 3: Operational chart of the developed viscometer

Methods previously used by researchers required that viscosity be calculated manually from measured parameters like: distance (cm), time (s), current (A) and voltage (V). This present study however overcame this hurdle as viscosity values were deduced directly from the sample and displayed on the screen. Figure 4 shows the image of the designed circuit.







2.2 **Viscometer Calibration**

The developed instrument was calibrated using water as a standard liquid with known viscosity value as provided in literature. It is necessary to compute the calibration constant K in order to ascertain the accuracy and efficiency of the developed system. The calibration of the viscometer was done using equation 1.1.



121

where η is the viscosity, K is the calibration constant, I is the current, V is the voltage and t is the time taken for the bob to complete an oscillation in a liquid. The microcontroller was used to measure the current, voltage and time which made the computation of the calibration constant easier. Figure 5 shows the complete circuit diagram of the developed system.

126



130

Figure 5: Circuit diagram of the developed viscometer

131132 The following results were obtained when water was used as the standard liquid;

- 133 Current = 0.249 A
- 134 Voltage = 12.022 V
- 135 Time (for to and fro) = 7.0 ms
- 136 Using equation 1.1, the calibration constant K is;
- 137

138

$K = \frac{1.0087}{0.249 \times 12.022 \times 0.007}$

- 139
- 140 $K = 48.137 \, mPa / VA$

141 The calculated calibration constant K was substituted into equation 1.0 and a suitable program was written and 142 embedded into the microcontroller so as to directly compute the viscosity η of any liquid under investigation using 143 equation 1.0.

144 145

146 3.0 RESULTS AND DISCUSSION

Five different liquid samples of known viscosities values were tested using the designed instrument under the same condition. The results obtained were compared with their corresponding standard values as shown in Table 1.0. The samples are water, olive oil, engine oil, gear oil and palm oil. The microcontroller performed the experimental measurement of period *t* during which the pendulum swings in the liquid sample, the supplied voltage *V* and current *l*, after which the measured parameters were used to compute and display the viscosity of the samples. The system was allowed to dry before another sample measurement was taken. At the end of the measurements, the values obtained were compared with known literature values of [12].

155

Table 1.0: Standard and Measured viscosities of various liquids at ambient temperature 157

Liquid	Time (m)	Voltage (V)	Current (A)	Literature Value	Instrument Value
				(mPa.s)	(mPa.s)
Water	7.028	12.03	0.25	1.0087	1.0237
Olive Oil	15.520	12.63	3.90	40.0000	37.0000
Palm oil	16.270	12.76	4.08	45.0000	41.0000
Engine Oil	60.780	13.83	9.32	390.0000	379.0000
Gear Oil	98.970	14.72	9.72	700.0000	686.0000

158 159



Figure 6.0: Plot between Measured viscosity values and literature values

164 165



Figure 7.0:Bar chart between the measured viscosity values and literature values

Statistical analysis was carried out and a Mean Absolute Percent Error (MAPE) of 2.87 % was obtained between the experimental and theoretical values. The plots of the measured and the standard values shows that developed system provides linear and reliable output as shown in Figures 6.0 and 7.0.

173 The correlation factor (R^2) of 0.999978 obtained also show a reliably good agreement between the measured values 174 with the standard data values. The instrument can be operated between 0°C and 100°C.

177 4. CONCLUSION

179 A cost effective locally made microcontroller-based digital vibrational viscometer has been developed to measure and 180 display viscosity of liquids. The calibration constant of the instrument was 48.137 mPa/VA which was calculated from 181 liquid of known (water) viscosity. The viscosity obtained by the developed instrument compared favorably well with the standard. It was observed that the Mean absolute percentage error of the measurement was found to be 2.87% which 182 183 translates to an accuracy of 97.13%. The instrument is highly reliable, user friendly, less expensive and portable. The 184 design and implementation of the instrument cost approximately \$50 while the prevailing market price of digital 185 viscometer is about \$400. Further investigation is still on-going using pendulum bob of various sizes to ascertain the 186 effect of bob sizes on viscosity. Also, the speed of the pendulum bob is being investigated and measured. A wider 187 output screen is being considered to accommodate a graphical user interface. A data logging section will also be 188 incorporated to create a means of keeping the data in a more refined and modern way for future purposes.

189 190

166 167

168 169

175 176

178

191 ACKNOWLEDGEMENTS192

The Authors wish to appreciate everyone who has contributed towards the success of this publication. However, it must be emphasized that this work receives no funding from any external source. The work was solely financed by the Authors.

REFERENCES

200		
200 201 202	[1]	K. Walters, W. M. James. Instrumentation reference book. Butterworth Heinemann Publishing Company, Oxford, U.K, 2010;(4): 69-75.
203		
204 205	[2]	R. Fox, A. T. McDonald, P. J. Pritchard. Fluid Mechanics, 8th Edition, Wiley New York. 2012;(8): 31-33
206 207 208	[3]	M. A. Eiteman, J. W. Godwin, Rheology of the Triglycerides Tricaprison, Triceprylin and Tricaprin of Diese Fuel. Transcaters ASAE, Emerald Publishing Limited, 1993; 503-507.
209 210	[4]	M. A. Rao. Rheology of Foods. Journal of Food Engineering. NY,USA. 1997:135-168

211	[5]	C. Klofutar, D. T. Rudan, K. Romin, Densities and thermal expansion Coefficients of Vegetables Oils, Acta
212 213		Chin.Slov 1997: 45-55
214 215 216 217	[6]	A. I. Abdulagatov, M. A. Magerramov, N. D. Azizor. Effect of Temperature, Pressure and Concentration on the Viscosity of Fruit Juices. Experimental And modeling, Physical and Chemical Properties Division, National Institute of Standards and Technology, U.S.A. 2011: (63).
218 219 220	[7]	R. Mostert, P. S. VanDer Gulik, H. R. Van Den Berg. The Working Equations of a Vibrating Wire Viscometer. Statistical and Theoretical Physics. Elsevier. 1989;56(3): 909-920.
221 222 223	[8]	M. Kandil. The Development of a Vibrating Wire Viscometer and a Microwave Cavity Resonator. Ph.D. dissertation, University of Canterbury, New Zealand. 2005.
224 225 226 227	[9]	T. Retsina, S. M. Richardson, W. A. Wakeham. The Theory of a Vibrating Rod Viscometer. Journal of Applied Scientific Research. 1987;3(2): 325-346
228 229 230 231	[10]	R. Won-kyu, O. Kenichi, Paul. Non-contact Technique for Measuring Surface Tension and Viscosity of Molten Materials using High Temperature Electrostatic Leviator. American Institute of Physics. Review of Scientific Instruments. 2013;(70): 2796-2801.
232 233 234	[11]	B. Y. Lawal, J. B. Dada, Ahmed-Ade. Construction of Automatic Temperature Controller for monitoring Heating and Cooling Systems. Journal of Sustainable Technolog. 2016; (7): 1-5.
235	[12]	Burkle Gmbh, Viscosity of Liquids Datasheet. 2012: 1-3.